SS-5710 Section 1

# **Specifications**

#### 1-1 GENERAL

The SS-5710 is an oscilloscope with a frequency bandwidth of DC to 60 MHz that can display 8 traces on 4 channels. The SS-5710 is useful in a wide range of applications for not only production lines and maintenance and service purposes but also for the research and development of a variety of electronic devices. The features of the SS-5710 are as follows:

- In addition to display of 8 traces on 4 channels, the SS-5710 has an ADD function for measuring the sum of two signals and CH 2 POLAR for measurement of the difference between two signals.
- Both CH 1 and CH 2 have a high deflection factor of 1 mV/div (in the x5 MAG function), which permits accurate measurement of lower voltages.
- The horizontal deflection system has sweep rates up to 5 nS/div (in the x10 MAG function) so that even high-speed phenomena can be measured with accuracy.
- The SS-5710 has delayed sweep, single sweep, ALT sweep, and X-Y operation functions, and a TV synchronizing signal separator circuit so that television and other composite video signal waveforms can be observed.

#### 1-2 ELECTRICAL SPECIFICATIONS

#### 1-2-1 Cathode-Ray Tube (CRT)

Shape Rectangular, 6 inches

Dispaly Area 8 div x10 div (1 div = 10 mm),

with internal illuminated

graticule of parallax-free type

Phosphor B31 (Standard)

Accelerating Voltage Approximately 15 kV

# 1-2-2 Vertical Deflection System

Modes CH 1, CH 2, ALT, CHOP,

ADD, QUAD (Quadruple) CHOP switching rate: 300

 $kHz~\pm40\%$ 

Channels 1 and 2

Rise Time

Deflection Factor 5 mV/div to 10 V/div, in

11 calibrated steps in a 1-2-5

sequence

Accuracy: ±2%

(at 10° C to 35° C)

 $\pm$  5%

(at  $-10^{\circ}$  C to  $50^{\circ}$  C)

5 mV/div to 25 V/div, continuously variable with the VARIABLE control

x5 MAG: 1 mV/div to 2 V/div in 11 calibrated steps

Accuracy: ±4%

(at 10° C to 35° C)

±8%

 $(at -10^{\circ} C to +50^{\circ} C)$ 

Frequency Response DC to 60 MHz, -3 dB

(5 mV/div to 0.2 V/div) DC to 20 MHz, -3 dB (1 mV/div, 2 mV/div in the x 5 MAG made

Notes

• 10° C to 35° C

• AC coupling: The lowest useable frequency is 4 Hz.

useable frequency is 4 Hz. 5.8 nsec (5 mV/div) or less

Fulse Response Overshoot: 5% or less

Sag (at 1 kHz):1.5% or less Other distortions: 5% or less (5 mV/div,  $10^{\circ}$  C to  $35^{\circ}$  C) Signal Delay

Delay cable supplied

Input Coupling

AC, DC, GND

Input RC

Direct:

1 M $\Omega$  ± 2%//32pF ±3pF

With probe:

10 M  $\Omega$  ±2%//15pF ± 2pF

Maximum Input Voltage

Direct:

250 V (DC +peak AC)

With probe:

600 V (DC +peak AC)

(Refer to the instruction manual for the probe for the maximum input voltage where

probe is used.)

Drift

0.5 div/hour (5 mV/div) or

2 div/hour (1 mV/div) 30 minutes after power is

turned on (Standard)

Common Mode Rejection Ratio

5 mV/div

40: 1 (1 kHz sine wave)

15: 1 (5 MHz sine wave)

Polarity Inversion

CH 2 only

Channels 3 and 4

Deflection Factor

0.1 V/div, 1 V/div, selectable

Accuracy: ±4%

(at 10 °C to 35 °C)

Frequency Response

DC to 50 MHz, -3 dB

Notes

•10°C to 35°C

·AC coupling: The lowest

usable frequency is 4 Hz.

Pulse Response

Overshoot: 10%

Sag (at 1 kHz): 2%

Other Distortions: 10%

Input Coupling

AC, DC

Input RC

Direct:

 $1 M\Omega \pm 3\% //32 pF \pm 3 pF$ 

With probe:

10 M  $\Omega \pm 2\%//15$  pF  $\pm 2$  pF

Maximum Input Voltage

Direc::

250 V (DC +peak AC)

With probe:

600V (DC +peak AC)

1-2-3 Triggering

A-Triggering

Triggering Mode

AUTO, NORM,

SING LE/RESET

Signal Source

CH 1, CH 2, CH 3, LINE,

NORM (External trigger can be used by selecting CH 3

with SOURCE switch.)

Coupling

AC, DC, HF REJ, LF REJ,

FIX, TV-H, TV-V

Slope

Possitive-going (+),

Negative-going (—)

Minimum Trigger Sensitivity

As shown in Table 1-1

Table 1-1

(at 10°C to 35°C)

Frequency Range	Sensitivity		
	CH 1, CF 2	CH 3, CH 4	
DC to 1 kHz	1 div	1.5 div	
1 kHz to 2 MHz	0.5 div	1 div	
2 MHz to 20 MHz	1 div	1.5 div	
20 MHz to 60 MHz	1.5 div	2 div	

Note

Fix: 1 div at 10 Hz to 2 MHz
 2 div at 2 MHz to 30 MHz
 Sine wave only

 TV-V, TV-H synchronizing signal level: 2.3 div or more on screen amplitude for a composite video signal

Section 1 Specifications

SS-5710		SS.	-57	1	0
---------	--	-----	-----	---	---

		composed of 7 parts video signal and 3 parts synchronizing signal  Trigger signals are attenuated in the following frequency ranges depending on coupling AC: 10 Hz or less  HF REJ: 10 kHz or higher LF REJ: 10 kHz or lower  AUTO sweep mode: The lowest useable frequency is	Hold-Off Time B-Sweep	±4% at 10 msec/div to 0.5 sec/div (at 10 °C to 35 °C) ±5% (at -10 °C to +50 °C) Accuracy II (Over any 2 of the center 8 divisions): ±5% (at -10 °C to +50 °C) Variable with the HOLDOFF control
-		50 Hz.	Delay	Continuous delay (RUNS AFTER DELAY), triggered
Weeke	<b>B-Triggering</b> Signal Sources	RUNS AFTER DELAY, CH  1, CH 2, CH 4 (External trigger can be used by selecting CH 4 with SOURCE switch.)	Sweep Rates	delay 50 nsec/div to 50 msec/div, in 19 calibrated steps in a 1-2-5 sequence Accuracy I (Over center 8 divisions):
or college	Coupling Slope	AC, DC, HF REJ, TV-H Positive-going (+), negative-going ()		$\pm 3\%$ (at 10 ° C to 35 ° C) $\pm 5\%$ (at -10 ° C to +50 ° C) Accuracy II (Over 2 of the
Name .	Minimum Trigger S	As showm in Table 1-1	Time Difference Management	center 8 divisions): $\pm 5\%$ (at $-10^{\circ}$ C to $+50^{\circ}$ C)
-		However, Sensitivity of 20 MHz to 60 MHz is 2 div at CH 1, CH 2.	Time Difference Measi	0.5 µsec/div to 5 sec/div Accuracy: ±2% of reading ±0.01 graduation (Minimum graduation of DELAY TIME
	1-2-4 Horizontal De	flection System	Delay Jitter	MULT dial) 1/20,000 or less
-	Modes	A, A INTEN, ALT, B (DLT'D), X-Y	Sweep Magnification	10 times (Maximum sweep rate: 5 nsec/
,-1880	A-Sweep Sweep Rates	50 nsec/div to 0.5 sec/div, in 22 calibrated steps in a		div) Accuracy I of magnified sweep rate (Over center 8 divisions)
And the second		1-2-5 sequence 50 nsec/div to 1.25 sec/div, continuously variable with		$\pm$ 5% at 50 nsec/div to 0.1 $\mu$ sec/div $\pm$ 4% of 0.2 $\mu$ sec/div to 0.5
-		the VARIABLE control Accuracy   (Over center 8 divisions):		sec/div (at 10° C to 35° C) Accuracy II of magnified sweep rate (Over any 2 of the
-		±3% at 50 nsec/div to 5 msec/div		center 8 divisions): ±10% at 50 nsec/div to 0.1 µsec/div

 $\pm 6\%$  at 0.2  $\mu$  sec/div to 0.5

sec/div (at 10° C to 35° C) (Except 25 nsec before and

after sweep)

Output Current 10 mA

Accuracy: ±2%

(at 10° C to 35° C)

±3%

(at  $-10^{\circ}$  C to  $50^{\circ}$ C)

#### 1-2-5 X-Y Operation

X Axis (Same as CH 1 except for the

following)

Deflection Factor Same as that of CH 1

Accuracy: ±5%

(at 10°C to 35°C)

±6%

(at  $-10^{\circ}$  C to  $+50^{\circ}$ C)

Frequency Response DC to 2 MHz, -3 dB

Y Axis same as CH 2

X-Y Phase Difference 3 or less (at DC to 50 kHz)

### 1-2-8 Power Supply

Voltage Range 100V ( 90 to 110 V)/

115V (103 to 128 V)/ 220V (195 to 242 V)/

230, 240V(207 to 264 V)/

AC

One of these voltage ranges

can be selected with voltage

selector plug

Frequency Range 50 to 440 Hz

Power Consumption Approximately 50 W

(at 100 VAC)

#### 1-2-6 Z-Axis System

Sensitivity 0.5 Vp-p

Polarity Positive decleases intensity,

negative incleases intensity

Frequency Range DC to 3 MHz

Input Resistance

5 k Ω±10%

Maximum Input Voltage

50 V (DC +peak AC)

## 1-3 PHYSICAL CHARACTERISTICS

Weight Approximately 8.5 kg

(without panel cover and

accessories bag)

Dimensions 320  $\pm$  2 (W) x 160  $\pm$  2 (H)

 $\times$  400  $\pm$  2 (L) (mm) See Figure 1-1.

#### 1-2-7 Calibrator

Waveform Square wave

Repetition Frequency 1 kHz

Accuracy: ±30%

(at 10 °C to 35 °C)

**Duty Ratio** 

40% to 60%

Output Voltage

0.3 V

Accuracy: ±1%

(at 10°C to 35°C)

±2%

(at -10 °C to +50 °C)

## 1-4 ENVIRONMENTAL CHARACTERISTICS

Operating Temperature -10 °C to -50 °C

Operating Humidity 40°C, 90% Relative Humidity

Storage Temperature -20°C to 70°C

Storage Humidity 70° C. 80% Relative Humidity

SS-5710 Section 1 Specifications

Altitude	Operating: 5,000 m maximum (atmospheric pressure 428	1-5	ACCESSORIES			
	mm Hg)		Power cord	1		
	Non-operating: 15,000 m		Probe (SS-0011)	2		
	maximum (atmospheric		Fuse (FSA-1)	2		
	pressure 87 mmHq)		Panel cover	1		
Vibration	From 10 Hz to 55 Hz and		Dust cover	1		
	back in 1 minute;		Instruction Manual	1		
	double amplitude 0.63 mm;		Accessories bag	1		
	for 15 minutes each in		For the method of	removing	the accessories	bag,
	vertical, horizontal, and longi-		refer to Figure 1-2.			
	tudinal directions for a total					
	of 45 minute					
Impact	One side is raised to an					
	elevation angle of 45° (10 cm					
	maximum), and let fall on a					
	piace of hard wood. Each side					
	is put to this test 3 times.					
Drop	A package ready for trans-					
	potation is dropped from a					
	height of 60 cm.					

Figure 1-1. Dimensional Diagram





Figure 1-2. Accessories Bag



Accessories bag mounting screw

When removing the accessories bag form the upper cover of the SS-5710, remove the four accessories bag mounting screws shown in Figure 1-2.

Use the same screws for mounting the accessories bag on the upper cover again.